ONERENT · QUICK USER GUIDE

Multifunction meter TESTO 435-4

A short guide to the multifunction meter: connecting probes, measuring and saving results.



1. Before you start

- Check that the kit contains everything listed in the hand-over record.
- Check the charge of the 3 AA batteries.
- Choose the probe you need (air velocity, humidity, CO_2, etc.) and plug it into the socket while the device is off.

2. Measuring

- Switch on the device — it recognises the probe and shows its readings.
- Hold the probe at the measuring point until the reading settles; with air velocity probes, follow the flow direction mark.
- Differential pressure is measured with the built-in sensor — connect the tubes to the “+” and “–” inputs; zero the sensor before measuring.

3. Saving

- Save the readings in memory — they can be grouped by location.
- Transfer the data to a computer with the testo software and a USB cable.

4. Tips

- To calculate air volume flow, enter the duct dimensions or area.
- Let humidity and CO_2 probes adapt to the room temperature for a few minutes.

5. Safety and care

- Do not touch hot-wire sensors and do not immerse probes in water.
- Connect and disconnect probes with the device switched off.
- If the equipment does not work properly, stop using it and call us: +371 26537465.

6. Returning

- Return the equipment to Slokas iela 52N clean and complete — with all accessories listed in the hand-over record.

Based on the manufacturer’s instructions (testo SE & Co. KGaA). Full manufacturer manual: www.testo.com.