ONERENT · QUICK USER GUIDE

Carbon monoxide (CO) meter TESTO 440 dP with CO probe (Bluetooth)

A short guide to the testo 440 dP kit with CO probe (Bluetooth): preparation, measuring carbon monoxide indoors and reports.



1. Before you start

- Check that the kit contains everything listed in the hand-over record.
- Check the batteries in the instrument (3 × AA) and in the probe (4 × AA).
- Switch on the probe and the instrument in fresh air (outdoors or in a ventilated room) — the probe connects via Bluetooth (up to 20 m).

2. Measuring

- In the instrument menu, choose the CO measurement and wait until the reading in fresh air settles close to 0 ppm.
- Enter the room to check and hold the probe at breathing height; near heating appliances, also measure close to them and at the door.
- Wait until the reading settles; a single reading can be saved with the button on the probe handle.

3. Saving

- Save the result as a report with the room or appliance name.
- Transfer reports to a computer with the USB cable.
- The built-in differential pressure sensor also measures pressure drop across filters: connect the tubes to “+” and “–” and zero the sensor before measuring.

4. Tips

- An increased CO level in a room points to incomplete combustion or poor flue gas removal — the appliance should be checked by a specialist.
- The probe measures CO in the ambient air, not in the flue gas — no flue opening is needed.

5. Safety and care

- If the reading rises quickly, leave the room at once, ventilate it and call a specialist or the emergency service (112).
- Do not put the probe into a flue gas stream and protect it from water and knocks; carry the probe and instrument in the case.
- If the equipment does not work properly, stop using it and call us: +371 26537465.

6. Returning

- Return the equipment to Slokas iela 52N clean and complete — with all accessories listed in the hand-over record.

Based on the manufacturer’s instructions (testo SE & Co. KGaA). Full manufacturer manual: www.testo.com.